



CERTIFICATE OF CALIBRATION

Page:1/1

Certificate No.	25090001	Date:23-09-2025	Job No.:	250900001
-----------------	----------	-----------------	----------	-----------

Part Description:	PRESSURE GAUGE			
Part No.:	G25151PS			
Serial Number:	****			
Range:	0-15 PSI /KPA			
Resolution:				
Accuracy:	+/- 1.6%			
		Dial Size:	2.5"(63MM)	
		Connection:	1/4" NPT	

Calibration Date:23-08-2023	Calibration Due:22-08-2024
-----------------------------	----------------------------

Environmental Conditions
Temperature: 25 °C Humidity: 50%

Master Equipment Details:					
Description	Serial No.	Model	Cal.Due	Certificate No.	Traceable Standard
Master Pressure gauge	E1501	323(0-15PSI)	30-12-2025	E150123	EN837-1

Test Results:

Master Reading Psi	Instruments Reading Psi	Error Psi
0	0	0.0
3	3.1	-0.1
9	9.11	-0.1
12	12.12	-0.1
15	15.01	0.0
Remarks:Maximum Error is -0.80%		
Result :	PASSED	